



Declaration of Conformity

Manufacturer: *ResMed Ltd*
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

European Representative: *ResMed (UK) Ltd*
96 Milton Park
Abingdon
Oxfordshire OX14 4RY
United Kingdom

Notified Body: *SGS United Kingdom Ltd*
Weston-Super-Mare
Unit 202b, Worle Parkway
Weston-Super-Mare, BS22 0WA
United Kingdom



Product: Swift FX

The Swift FX channels airflow noninvasively to a patient from a positive airway pressure device such as a continuous positive airway pressure (CPAP) or bilevel system.

The Swift FX is:

- to be used by adult patients (>30 kg) for whom positive airway pressure has been prescribed
- intended for single-patient re-use in the home and multipatient re-use in the hospital/institutional environment.

Standards Applied: ISO 14971:2007; ISO 17510-2:2007;
ISO 10993-1:2003; ISO 5356-1:2004 (sections 5.1 and 5.2);
ISO 17664:2004.

Classification: IIa

Conformity

Assessment Route: Annex II (excluding Section 4), 93/42/EEC

We herewith declare that the above mentioned products meet the transposition into national law of the provisions of Council Directive 93/42/EEC including the MDD amendment (2007/47/EC), for medical devices. All supporting documentation is retained at the premises of the manufacturer.

Sydney,

Steven Lubke

STEVEN LUBKE

6 AUGUST, '09.

DIRECTOR - REGULATORY AFFAIRS

for
Dr. Lionel King
Vice President Global Quality Assurance and Regulatory Affairs
ResMed Ltd

EC107